



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

***Calip-One Weighing and Inventory Specialists, Inc.
57 Tournament Drive South
Hawthorn Woods, IL 60047***

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

*This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):*

***Calibration of Scales and Balances
(As detailed in the supplement)***

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:

August 04, 2006

Issue Date:

November 11, 2009

Expiration Date:

November 10, 2011

Accreditation No.:

59075

Certificate No.:

L09-109

Page No.:

Page 1 of 2



Certificate of Accreditation: Supplement

Calip-One Weighing and Inventory Specialists, Inc.
57 Tournament Drive South
Hawthorn Woods, IL 60047

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED QUANTITY, INSTRUMENT OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Micro and Semi-Micro Analytical Balance	0.001 g to 160 g *Possible Multi – Range	$(1.16 \times 10^{-3} + 1 \times 10^{-6} \text{wt}) \text{ g}$	Class I Devices HB 44 Test Methods Class 1 Weights
Electronic Balances Top Loader Type	0.01 g to 64 000 g *Possible Multi – Range	$(1.15 \times 10^{-2} + 4 \times 10^{-6} \text{wt}) \text{ g}$	Class I/II Devices HB 44 Test Methods Class 1 Weights
	0.01 g to 64 000 g *Possible Multi – Range	$(1.15 \times 10^{-2} + 6 \times 10^{-6} \text{wt}) \text{ g}$	Class I/II Devices HB 44 Test Methods Class 2 Weights
	0.01 g to 64 000 g *Possible Multi – Range	$(1.15 \times 10^{-2} + 2.3 \times 10^{-5} \text{wt}) \text{ g}$	Class I/II Devices HB 44 Test Methods Class 4 Weights
	0.01 g to 64 000 g *Possible Multi – Range	$(1.15 \times 10^{-2} + 1.16 \times 10^{-4} \text{wt}) \text{ g}$	Class I/II Devices HB 44 Test Methods Class F Weights
Dial and Beam Scales Bench Scales Floor Scales	0.454 kg to 2 267.96 kg (1 lb to 5 000 lb)	$(5.24 \times 10^{-1} + 2.8 \times 10^{-5} \text{wt}) \text{ kg}$ $[(1.16 + 2.8 \times 10^{-5} \text{wt}) \text{ lb}]$	Class III/III L Devices HB 44 Test Methods Class F Weights
Floor Scales Hopper Scales Tank Scales	22.68 kg to 99 790.32 kg (50 lb to 220 000 lb)	86.183 kg (190 lb)	Class III/III L Devices HB 44 Test Methods Class F Weights

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent method information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.
2. * Possible Multi-Range – Multi Range balances are common for these type devices, dependent on client's requirements and usage, the testing and calibrations will be performed and is written to cover all ranges.
3. The lower end of scale calibration ranges represents the smallest weight a laboratory would use when performing a calibration. The scale may in fact be capable of weighing smaller values.
4. The term wt represents weight in pounds or grams (including SI multiple and sub multiple units) appropriate to the uncertainty statement.