



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

***Accurate Superior Scale Co., Inc.
5404 Jedmed Ct
Saint Louis, MO 63129-2221***

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

***Calibration of Scales and Balances
(As detailed in the supplement)***

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

<i>Initial Accreditation Date:</i>	<i>Issue Date:</i>	<i>Revision Date:</i>	<i>Expiration Date:</i>
August 28, 2002	July 07, 2008	January 15, 2009	July 06, 2010

<i>Accreditation No.:</i>	<i>Certificate No.:</i>	<i>Page No.:</i>
59049	L08-48-R1	Page 1 of 2



Certificate of Accreditation: Supplement

Accurate Superior Scale Co., Inc.
5404 Jedmed Ct
Saint Louis, MO 63129-2221

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	Remarks
Precision Balance Class I	5 g to 220 g	$(2.09 \times 10^{-2} + 4.00 \times 10^{-6} \text{ wt}) \text{ mg}$	Handbook 44 Class S1
	220 g to 500 g	$(4.34 \times 10^{-1} + 2.00 \times 10^{-6} \text{ wt}) \text{ mg}$	
	220 g to 1 000 g	$(11.33 + 1.00 \times 10^{-6} \text{ wt}) \text{ mg}$	
Balances & Bench Scales Class II	220 g to 1 000 g	$(11.36 + 1.00 \times 10^{-6} \text{ wt}) \text{ mg}$	Handbook 44 Class S1
Bench Scales Class III	907.18 kg to 4 535.90 g	$(4.82 \times 10^{-1} + 5.70 \times 10^{-5} \text{ wt}) \text{ g}$	Handbook 44 Class F
	11 339.75 kg to 22 679.51 g	$(2.09 + 7.30 \times 10^{-5} \text{ wt}) \text{ g}$	
	23 133.09 kg to 45 359 g	$(4.30 + 6.80 \times 10^{-5} \text{ wt}) \text{ g}$	
Bench, Floor, Crane, Lift Truck, Tank / Hopper, Class III	45.81 kg to 90.78 kg	$(8.46 \times 10^{-3} + 6.90 \times 10^{-5} \text{ wt}) \text{ g}$	Handbook 44 Class F
	91.4 kg to 224.3 kg	$(2.18 \times 10^{-2} + 6.90 \times 10^{-5} \text{ wt}) \text{ g}$	
	227.7 kg to 454.5 kg	$(4.32 \times 10^{-2} + 6.70 \times 10^{-5} \text{ wt}) \text{ kg}$	
	455 kg to 2 272.7 kg	$(2.39 \times 10^{-1} + 5.90 \times 10^{-5} \text{ wt}) \text{ kg}$	
	2 273.2 kg to 4 545.9 kg	$(4.34 \times 10^{-1} + 6.70 \times 10^{-5} \text{ wt}) \text{ kg}$	
Vehicle Scales, Tank/Hopper, Crane Scales Class III	4 545.9 kg to 9 090.9 kg	$(10.45 + 9.00 \times 10^{-6} \text{ wt}) \text{ kg}$	Handbook 44 Class F
	9 090.9 kg to 22 727.7 kg	$(26.18 + 8.00 \times 10^{-6} \text{ wt}) \text{ kg}$	
	22 727.7 kg to 45 454.5 kg	$(52.23 + 9.00 \times 10^{-6} \text{ wt}) \text{ kg}$	
	45 455 kg to 90 909.1 kg	$(104.47 + 9.00 \times 10^{-6} \text{ wt}) \text{ kg}$	
	90 909.5 kg to 181 818.2 kg	$(261.45 + 7.00 \times 10^{-6} \text{ wt}) \text{ kg}$	

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.
2. The term wt represents weight in pounds or grams (including SI multiple and submultiples units) appropriate to the uncertainty statement.