

# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:*

***American Scale Corporation  
3540 Bashford Ave  
Louisville, KY 40218***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

***ISO/IEC 17025: 2005***

*This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-FLAC-IAF Communiqué dated January 2009):*

***Laboratory and Laboratory Controlled Field Calibrations of Weighing Devices  
(As detailed in the supplement)***

*Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.*

For PJLA:

*The validity of this certificate is mandated through ongoing surveillance.*

---

Tracy Szerszen  
President/Operations Manager

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
26555 Evergreen, Suite 1325  
Southfield, Michigan 48076

*Initial Accreditation Date:*

January 09, 2004

*Issue Date:*

February 03, 2008

*Revision Date:*

January 15, 2009

*Expiration Date:*

February 02, 2010

*Accreditation No.*

59225

*Certificate No.*

L07-116-R1

*Page No.*

Page 1 of 2

# *Certificate of Accreditation: Supplement*

**American Scale Corporation**  
3540 Bashford Ave  
Louisville, KY 40218

*Accreditation is granted to this facility to perform the following calibrations:*

## Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	REMARKS
Lab Balances Class 1 devices	0.001 g to 100 g	$(1.16 \times 10^{-3} + 1.00 \times 10^{-6} \text{wt}) \text{ g}$	Method HB44 NISTIR 6919 Class 1 ASTM Weight Standards
Lab Balances Class 2 & 4 devices	100 g to 2 000 g	$(1.13 \times 10^{-2} + 3.00 \times 10^{-6} \text{wt}) \text{ g}$	Method HB44 NISTIR 6919 Class 2 ASTM Weight Standards
Industrial Scales, Balances, Tank & Hopper Scales Class F devices	1 g to 500 g	$(1.15 \times 10^{-2} + 1.41 \times 10^{-4} \text{wt}) \text{ g}$	Method HB44 HB 105 Class F Weight Standards
Industrial Scales, Balances, Tank & Hopper Scales Class F devices	0.001 lb to 10 lb	$(1.16 \times 10^{-3} + 4.80 \times 10^{-5} \text{wt}) \text{ lb}$	Method HB44 HB 105 Class F Weight Standards
Industrial Scales, Bench, Crane, Floor, Livestock, Lift Truck, Tank & Hopper Scales Class F devices	10 lb to 20 000 lb	$(2.31 + 4.80 \times 10^{-5} \text{wt}) \text{ lb}$	Method HB44 HB 105 Class F Weight Standards
Heavy Capacity Industrial Scales, Tank & Hopper Scales Class F devices	20 000 lb to 100 000 lb	$(10.64 + 5.70 \times 10^{-5} \text{wt}) \text{ lb}$	Method HB44 HB 105 Class F Weight Standards
Vehicle Scales Class F devices	1 000 lb to 200 000 lb	$(16.27 + 6.10 \times 10^{-5} \text{wt}) \text{ lb}$	Method HB44 HB 105 Class F Weight Standards

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.
2. The term wt represents weight in pounds or grams (including SI multiple and submultiples units) appropriate to the uncertainty statement.