

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

***C & M Scale Company
7241 West Roosevelt Road
Forest Park, Illinois 60130***

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-FLAC-IAF Communiqué dated January 2009):

***Calibration of Scales and Precision Balances
(As detailed in the supplement)***

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:

August 04, 2004

Issue Date:

October 18, 2008

Revision Date:

January 15, 2009

Expiration Date:

October 17, 2010

Accreditation No.

59304

Certificate No.

L08-86-R1

Page No.

Page 1 of 2

Certificate of Accreditation: Supplement

C & M Scale Company
7241 West Roosevelt Road
Forest Park, Illinois 60130

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Analytical Balances	0.01 g to 200 g	$(1.16 \times 10^{-3} + 1.00 \times 10^{-6} \text{wt})$ g	Class 1
Lab Balance	201 g to 5.0 kg	$(1.12 \times 10^{-2} + 4.00 \times 10^{-6} \text{wt})$ g	Class 2
Bench Scales	0.1 lb to 300 lb	$(1.15 \times 10^{-2} + 8.40 \times 10^{-5} \text{wt})$ lb	Class F 50 lb Weights
Platform Scales	301 lb to 3 000 lb	$(9.32 \times 10^{-2} + 9.10 \times 10^{-5} \text{wt})$ lb	Class F 500 lb Weights
Floor Scales	3 001 lb to 40 000 lb	$(9.14 \times 10^{-1} + 9.70 \times 10^{-5} \text{wt})$ lb	Class F 1 000 lb Weights
High Capacity Scales	40 001 lb to 200 000 lb	$(56.84 + 2.70 \times 10^{-5} \text{wt})$ lb	Class F 1 000 lb Weights

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.
2. The term wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.