



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

***Central Illinois Scale Company, Inc.
402 South L. Street, Tilton, IL 61833
2560 Parkway Court, Decatur, IL 62526
6000 Washington Street, Bartonville, IL 61607***

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

***Laboratory and Field Calibration of Scales and Weighing Systems
(As detailed in the supplement)***

Decatur and Bartonville are Remote Calibration Facilities Only

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

A handwritten signature in black ink, appearing to read 'Tracy Szerszen', is written over a horizontal line.

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:
February 23, 2007

Issue Date:
April 11, 2008

Revision Date:
December 18, 2009

Expiration Date:
April 10, 2010

Accreditation No.:
59078

Certificate No.:
L08-15-R1

Page No.:
Page 1 of 2



Certificate of Accreditation: Supplement

Central Illinois Scale Company, Inc.
402 South L. Street, Tilton, IL 61833
2560 Parkway Court, Decatur, IL 62526
6000 Washington Street, Bartonville, IL 61607

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Micro, Semi-Micro, Analytical Balances	2 g to 500 g	$(1.15 \times 10^{-2} + 1.00 \times 10^{-6} \text{ wt}) \text{ g}$	Class 1 ASTM Weights NIST Handbook 44 NISTIR 6919
Top Load Balances	500 g to 64 000 g	$(3.47 \times 10^{-4} + 2.00 \times 10^{-6} \text{ wt}) \text{ g}$	Class 1 ASTM Weights NIST Handbook 44 NISTIR 6919/ HB105-1
Moisture Analyzers	2 g to 500 g	$(1.15 \times 10^{-2} + 1.00 \times 10^{-6} \text{ wt}) \text{ g}$	Class 1 ASTM Weights NIST Handbook 44 NISTIR 6919
Mass Comparators	2 g to 5 000 kg	$(5.77 \times 10^{-7} + 2.00 \times 10^{-6} \text{ wt}) \text{ g}$	Class 1 ASTM Weights NIST Handbook 44 NISTIR 6919/ HB105-1
Bench Scales	2 lb to 1 000 lb	$(1.13 \times 10^{-2} + 1.05 \times 10^{-4} \text{ wt}) \text{ lb}$	Class F ASTM Weights NIST Handbook 44 HB105
Floor Scales	500 lb to 60 000 lb	$(1.14 + 3.50 \times 10^{-5} \text{ wt}) \text{ lb}$	
Hopper & Tank Scales	50 lb to 500 000 lb	$(11.55 + 9.50 \times 10^{-5} \text{ wt}) \text{ lb}$	
Mechanical Beam, Dial Head, Medical, Crane Livestock, Axle, Straight Weigh, and other Scales	50 lb to 500 000 lb	$(1.15 + 1.14 \times 10^{-4} \text{ wt}) \text{ lb}$	
Truck and Railroad Track Scales	25 000 lb to 500 000 lb	$(21.23 + 8.20 \times 10^{-5} \text{ wt}) \text{ lb}$	

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.
2. The term wt represents weight in pounds or grams (including SI multiple and submultiples units) appropriate to the uncertainty statement.