

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

**Century Scale Company
23731 Dequindre Rd
Hazel Park, MI 48030**

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-FLAC-IAF Communiqué dated January 2009):

**Force Calibration
(As detailed in the supplement)**

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:
May 10, 2004

Issue Date:
June 26, 2008

Revision Date:
January 16, 2009

Expiration Date:
June 25, 2010

Accreditation No.
59235

Certificate No.
L08-39-R1

Page No.
Page 1 of 2

Certificate of Accreditation: Supplement

Century Scale Company
23731 Dequindre Rd
Hazel Park, MI 48030

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Force Testing Machine	0 kg to 450 kg	0.9 kg	Class F Weights
	0 kg to 13 500 kg	27 kg	

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.