

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

**Charles S. Freeman Company, Inc.
3755 Harlem Road Buffalo, NY 14215
6700 Joy Road East Syracuse, NY 13057**

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-FLAC-IAF Communiqué dated January 2009):

**Buffalo Facility: Calibration of weighing devices (bench scales, platform scales, truck scales, railroad scales, count/weigh scales, crane scales, price computing scales, track scales, laboratory balances, precision laboratory balances, industrial balances), force measuring devices and mass (standard test weights)
Syracuse Facility: Calibration of weighing devices only
(As detailed in the supplement)**

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:
February 17, 2002

Issue Date:
April 04, 2008

Revision Date:
January 16, 2009

Expiration Date:
April 03, 2010

Accreditation No.
59083

Certificate No.
L08-12-R1

Page No.
Page 1 of 2

Certificate of Accreditation: Supplement

Charles S. Freeman Company, Inc.
3755 Harlem Road Buffalo, NY 14215
6700 Joy Road East Syracuse, NY 13057

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Balances	0.01 kg to 12 kg	0.06 g	Class F NIST HB 44
High Resolution Scales	12 kg to 300 kg	7.5 g	Class F NIST HB 44
Scales	0.454 kg to 90 718 kg (1 lb to 200 000 lb)	9.1 kg 20 lb	Class F NIST HB 44

1. Syracuse facility accreditation is for calibration of weighing devices only
2. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.