

# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:*

***Derlein Scale, Inc.  
1639 Edgewood Rd.SW  
Cedar Rapids, IA 52404***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

***ISO/IEC 17025: 2005***

*This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-FLAC-IAF Communiqué dated January 2009):*

***Laboratory and Field Calibration of Commercial, Industrial and  
Vehicle Scales and Weighing Devices to NIST Handbook 44  
(As detailed in the supplement)***

*Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.*

For PJLA:

*The validity of this certificate is mandated through ongoing surveillance.*

---

Tracy Szerszen  
President/Operations Manager

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
26555 Evergreen, Suite 1325  
Southfield, Michigan 48076

*Initial Accreditation Date:*

May 10, 2008

*Issue Date:*

May 10, 2008

*Revision Date:*

January 16, 2009

*Expiration Date:*

May 09, 2010

*Accreditation No.*

59429

*Certificate No.*

L08-27-R1

*Page No.*

Page 1 of 2

# Certificate of Accreditation: Supplement

**Derlein Scale, Inc.**  
1639 Edgewood Rd.SW  
Cedar Rapids, IA 52404

*Accreditation is granted to this facility to perform the following calibrations:*

## Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	REMARKS
Heavy Capacity Industrial Scales	680.39 kg to 2 267.97 kg (1 500 lb to 5 000 lb)	1.36 kg (3 lb)	Class F Weights Handbook 44 Test Method
	2 268.42 kg to 4 535.93 kg (5 001 lb to 10 000 lb)	2.73 kg (6 lb)	
Industrial Scales Livestock, Crane, Hopper, Bench and Floor Scales	22.68 kg to 4 535.93 kg (50 lb to 10 000 lb)	2.73 kg (6 lb)	Class F Weights Handbook 44 Test Method
Light Capacity Scales and Balances	0.89 g to 22.68 kg (0.001 9 lb to 50 lb)	0.28 kg (0.060 lb)	Class F Weights Handbook 44 Test Method
Precision Laboratory Scales and Balances	0.010 g to 5 000 g	6.1 g	Class F Weights Handbook 44 Test Method

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services