

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

**Geo Nable Instrument Services
1199 Saxonburg Blvd.
Glenshaw, PA 15116**

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-FLAC-IAF Communiqué dated January 2009):

**Calibration of Optical Measuring Devices
(As detailed in the supplement)**

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:

July 06, 2006

Issue Date:

October 06, 2008

Revision Date:

January 22, 2009

Expiration Date:

October 05, 2010

Accreditation No.

59401

Certificate No.

L08-82-R1

Page No.

Page 1 of 2

Certificate of Accreditation: Supplement

Geo Nable Instrument Services
1199 Saxonburg Blvd.
Glenshaw, PA 15116

Accreditation is granted to this facility to perform the following calibrations:

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Magnification Accuracy of Optical Comparators	0 mm to 50 mm stage scale (10x to 100x Magnification)	0.004 1 mm	Procedure: P10B 300mm Glass Scale 50 mm Crosshair Scales Precision Ball and Pin Gages
	0 mm to 100 mm stage scale (10x to 100x Magnification)		
Magnification Accuracy of Profile Projector	0 mm to 50 mm stage scale (10x to 100x Magnification)	0.011 mm	Procedure: P10C 300mm Glass Scale 50 mm Crosshair Scales
	0 mm to 100 mm stage scale (10x to 100x Magnification)		
Accuracy of Measuring Microscope	0 mm to 50 mm stage scale (10x to 30x Magnification)	0.002 6 mm	Procedure: P10D 50 mm Stage Scale 50 mm Crosshair Scale
	0 mm to 100 mm stage scale (10x to 30x Magnification)		
Eyepiece Micrometer Microscope	0 mm to 50 mm stage scale (50x to 1 000x Magnification)	0.000 98 mm	Procedure: P10F & P10J Stage Micrometer and 50 mm Crosshair Scale
	0 mm to 100 mm stage scale (50x to 1 000x Magnification)		
Digital Camera Measuring System on Microscope	50x to 1 000x Optical Magnification of Software Generated μ meter Scale	1.1 μ m	Procedure: P10I & P10J Stage Micrometer and Glass Scale

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.