

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

***AI Calibration Laboratory, Inc.
3474 West 84th St. Bay 104
Hialeah, FL 33018***

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-FLAC-IAF Communiqué dated January 2009):

***Laboratory and Field Calibration of Dimensional, Pressure, Torque,
Thermodynamic and Electrical Measuring Devices
(As detailed in the supplement)***

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:

March 06, 2006

Issue Date:

April 09, 2008

Revision Date:

February 20, 2009

Expiration Date:

April 08, 2010

Accreditation No.

59381

Certificate No.

L08-16-R1

Page No.

Page 1 of 3

Certificate of Accreditation: Supplement

A1 Calibration Laboratory, Inc.
3474 West 84th St. Bay 104
Hialeah, FL 33018

Accreditation is granted to this facility to perform the following calibrations:

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Micrometers	0 mm to 101.6 mm	0.002 9 mm	

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Pressure Gages and Transducers	689.48 kPa to 6 894.75 kPa	9.3 kPa	
	34.473 MPa to 51.710 MPa	29 kPa	

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
DC Voltage	32 mV to 329 mV	88 μ V	
AC Voltage @ 1 kHz	32 mV to 329 mV	1 mV	
Thermocouple Type K Simulate / Measure	73.15 K to 173.15 K	0.33 K	
	173.15 K to 248.15 K	0.18 K	
	248.15 K to 393.15 K	0.16 K	
	393.15 K to 1 273.15 K	0.26 K	
	1 273.15 K to 1 645.15 K	0.40 K	
Thermocouple Type J Simulate / Measure	63.15 K to 173.15 K	0.27 K	
	173.15 K to 243.15 K	0.16 K	
	243.15 K to 423.15 K	0.14 K	
	423.15 K to 1 033.15 K	0.17 K	
	1 033.15 K to 1 473.15 K	0.23 K	
Thermocouple Type E Simulate / Measure	23.15 K to 173.15 K	0.50 K	
	173.15 K to 248.15 K	0.16 K	
	248.15 K to 623.15 K	0.14 K	
	623.15 K to 923.15 K	0.16 K	
	923.15 K to 1 273.15 K	0.21 K	

Certificate of Accreditation: Supplement

A1 Calibration Laboratory, Inc.
3474 West 84th St. Bay 104
Hialeah, FL 33018

Accreditation is granted to this facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Thermocouple Type T Simulate / Measure	23.15 K to 123.15 K	0.63 K	
	123.15 K to 273.15 K	0.24 K	
	273.15 K to 393.15 K	0.16 K	
	393.15 K to 673.15 K	0.14 K	

1. This laboratory offers commercial calibration services.
2. "Best uncertainty" is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards of nearly ideal measuring equipment. Best uncertainties represent expanded uncertainties expressed at approximately the 95% level of confidence, usually using a coverage factor of $k = 2$. The best uncertainty of a specific calibration performed by the laboratory may be greater than the best uncertainty due to the behavior of the customer's device, to the environment (of the calibration is performed in the field) and to influences from the circumstances of the specific calibration. Where ranges are not specified, the best measurement uncertainty stated is for the cardinal points only.
3. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.