

# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:*

***Dimensional & Calibration Laboratories, Inc.  
70 1st St  
Hackensack, NJ 07601-2023***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

***ISO/IEC 17025: 2005***

*This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):*

***Dimensional Inspection of Customer Supplied Product  
(As detailed in the supplement)***

*Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.*

For PJLA:

*The validity of this certificate is mandated through ongoing surveillance.*

---

Tracy Szerszen  
President/Operations Manager  
Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
26555 Evergreen, Suite 1325  
Southfield, Michigan 48076

<i>Initial Accreditation Date:</i> June 07, 2004	<i>Issue Date:</i> December 10, 2008	<i>Revision Date:</i> February 26, 2009	<i>Expiration Date:</i> December 09, 2010
<i>Accreditation No.</i> 59242	<i>Certificate No.</i> L08-98-1 R2	<i>Page No.</i> Page 1 of 2	

# *Certificate of Accreditation: Supplement*

**Dimensional & Calibration Laboratories, Inc.**

70 1st St

Hackensack, NJ 07601-2023

*Accreditation is granted to this facility to perform the following testing:*

FIELD OF TEST	ITEMS, MATERIAL OR PRODUCT TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Dimensional Inspection	Customer Supplied Products	Length 81.28 cm x 60.96 cm x 43.18 cm (32 in x 24 in x 17 in)	Mitutoyo FJ805 CMM	0.26 $\mu\text{m}$ (10 $\mu\text{in}$ )
		Length 30.48 cm x 20.32 cm (12 in x 8 in)	S-T Industries 24 in Optical Comparator	1.27 $\mu\text{m}$ (50 $\mu\text{in}$ )
		Length 20.32 cm x 10.16 cm x 12.7 cm (8 in x 4 in x 5 in)	S-T Industries Video Measuring Machine	0.26 $\mu\text{m}$ (10 $\mu\text{in}$ )
		Length 0 cm to 101.6 cm (0 to 40 in) ID/OD	Fowler Trimos V1000	1.27 $\mu\text{m}$ (50 $\mu\text{in}$ )
		Length 0 cm to 101.6 cm (0 to 40 in) ID/OD	Fowler Trimos THV	0.026 $\mu\text{m}$ (1 $\mu\text{in}$ )
		Contour 10.16 cm x 5.08 cm (4 in x 2 in)	Mitutoyo Contracer CHB	0.26 $\mu\text{m}$ (10 $\mu\text{in}$ )
		Roundness To 30.48 cm (12 in) diameter	Mitutoyo Roundtest RA211	0.026 $\mu\text{m}$ (1 $\mu\text{in}$ )
		Surface Finish 0 $\mu\text{m}$ to 812.8 $\mu\text{m}$ (0 $\mu\text{in}$ to 32 000 $\mu\text{in}$ ) Ra	Mitutoyo Surftest Sj401	0.000 26 $\mu\text{m}$ (0.01 $\mu\text{in}$ )