

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

**Georgia Scale Company
313 Commerce Drive
Martinez, GA 30907**

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-PLAC-IAF Communiqué dated January 2009):

**Calibration of Industrial and Laboratory Weighing Systems
(As detailed in the supplement)**

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:

July 15, 2003

Issue Date:

May 10, 2009

Expiration Date:

May 09, 2011

Accreditation No.:

59247

Certificate No.:

L09-49

Page No.:

Page 1 of 2

Certificate of Accreditation: Supplement

Georgia Scale Company
313 Commerce Drive
Martinez, GA 30907

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Analytical Balances	0.5 g to 20 kg with 1 g resolution	1.2 g	Class 1 Weights (ASTM)
Bench Scales	0.1 g to 5 kg with 0.5 g resolution	1.4 g	Class F Weights (NIST)
	0.5 g to 25 kg with 2 g resolution	3.4 g	
Floor Scales	0.1 kg to 453.592 kg with 0.1 kg resolution (0.22 lb to 1 000 lb with 0.2 lb resolution)	100 g (0.22 lb)	
	0.2 kg to 907.185 kg with 0.2 kg resolution (.44 lb to 2 000 lb with 0.5 lb resolution)	200 g (0.44 lb)	
	0.5 kg to 2 267.962 kg with 0.5 kg resolution (1.10 lb to 5 000 lb with 1 lb resolution)	460 g (1.2 lb)	
	1 kg to 4 535.924 kg with 0.9 kg resolution (2.2 lb to 10 000 lb with 2 lb resolution)	910 g (2 lb)	
Tank and Hopper Scales	10 kg to 31 751 kg with 9 kg resolution (22 lb to 70 000 lb with 20 lb resolution)	14 kg (31 lb)	
	10 kg to 45 359.237 kg with 9 kg resolution (22 lb to 100 000 lb with 20 lb resolution)	23 kg (51 lb)	
Truck Scales	10 kg to 90 718.474 kg with 9 kg resolution (22 lb to 200 000 lb with 20 lb resolution)	28 kg (62 lb)	
Rail Scales	25 kg to 181 436.948 kg with 22 kg resolution (56 lb to 400 000 lb with 50 lb resolution)	48 kg (110 lb)	

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

**Georgia Scale Company
313 Commerce Drive
Martinez, GA 30907**

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

**Calibration of Industrial and Laboratory Weighing Systems
(As detailed in the supplement)**

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:

July 15, 2003

Issue Date:

May 10, 2009

Expiration Date:

May 09, 2011

Accreditation No.:

59247

Certificate No.:

L09-49

Page No.:

Page 1 of 2

Certificate of Accreditation: Supplement

Georgia Scale Company
313 Commerce Drive
Martinez, GA 30907

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Analytical Balances	0.5 g to 20 kg with 1 g resolution	1.2 g	Class 1 Weights (ASTM)
Bench Scales	0.1 g to 5 kg with 0.5 g resolution	1.4 g	Class F Weights (NIST)
	0.5 g to 25 kg with 2 g resolution	3.4 g	
Floor Scales	0.1 kg to 453.592 kg with 0.1 kg resolution (0.22 lb to 1 000 lb with 0.2 lb resolution)	100 g (0.22 lb)	
	0.2 kg to 907.185 kg with 0.2 kg resolution (.44 lb to 2 000 lb with 0.5 lb resolution)	200 g (0.44 lb)	
	0.5 kg to 2 267.962 kg with 0.5 kg resolution (1.10 lb to 5 000 lb with 1 lb resolution)	460 g (1.2 lb)	
	1 kg to 4 535.924 kg with 0.9 kg resolution (2.2 lb to 10 000 lb with 2 lb resolution)	910 g (2 lb)	
Tank and Hopper Scales	10 kg to 31 751 kg with 9 kg resolution (22 lb to 70 000 lb with 20 lb resolution)	14 kg (31 lb)	
	10 kg to 45 359.237 kg with 9 kg resolution (22 lb to 100 000 lb with 20 lb resolution)	23 kg (51 lb)	
Truck Scales	10 kg to 90 718.474 kg with 9 kg resolution (22 lb to 200 000 lb with 20 lb resolution)	28 kg (62 lb)	
Rail Scales	25 kg to 181 436.948 kg with 22 kg resolution (56 lb to 400 000 lb with 50 lb resolution)	48 kg (110 lb)	

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

**Georgia Scale Company
313 Commerce Drive
Martinez, GA 30907**

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-PLAC-IAF Communiqué dated January 2009):

**Calibration of Industrial and Laboratory Weighing Systems
(As detailed in the supplement)**

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:

July 15, 2003

Issue Date:

May 10, 2009

Expiration Date:

May 09, 2011

Accreditation No.:

59247

Certificate No.:

L09-49

Page No.:

Page 1 of 2

Certificate of Accreditation: Supplement

Georgia Scale Company
313 Commerce Drive
Martinez, GA 30907

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Analytical Balances	0.5 g to 20 kg with 1 g resolution	1.2 g	Class 1 Weights (ASTM)
Bench Scales	0.1 g to 5 kg with 0.5 g resolution	1.4 g	Class F Weights (NIST)
	0.5 g to 25 kg with 2 g resolution	3.4 g	
Floor Scales	0.1 kg to 453.592 kg with 0.1 kg resolution (0.22 lb to 1 000 lb with 0.2 lb resolution)	100 g (0.22 lb)	
	0.2 kg to 907.185 kg with 0.2 kg resolution (.44 lb to 2 000 lb with 0.5 lb resolution)	200 g (0.44 lb)	
	0.5 kg to 2 267.962 kg with 0.5 kg resolution (1.10 lb to 5 000 lb with 1 lb resolution)	460 g (1.2 lb)	
	1 kg to 4 535.924 kg with 0.9 kg resolution (2.2 lb to 10 000 lb with 2 lb resolution)	910 g (2 lb)	
Tank and Hopper Scales	10 kg to 31 751 kg with 9 kg resolution (22 lb to 70 000 lb with 20 lb resolution)	14 kg (31 lb)	
	10 kg to 45 359.237 kg with 9 kg resolution (22 lb to 100 000 lb with 20 lb resolution)	23 kg (51 lb)	
Truck Scales	10 kg to 90 718.474 kg with 9 kg resolution (22 lb to 200 000 lb with 20 lb resolution)	28 kg (62 lb)	
Rail Scales	25 kg to 181 436.948 kg with 22 kg resolution (56 lb to 400 000 lb with 50 lb resolution)	48 kg (110 lb)	

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.