

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

**AAA Scale Co. Inc.
3212 Lattice Road
Wilson, NC 27895**

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-FLAC-IAF Communiqué dated January 2009):

**Calibration of Scales and Balances
(As detailed in the supplement)**

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:
January 28, 2003

Accreditation No.:
59222

Issue Date:
May 17, 2009

Certificate No.:
L09-59

Expiration Date:
May 16, 2011

Page No.:
Page 1 of 2

Certificate of Accreditation: Supplement

AAA Scale Co. Inc.
3212 Lattice Road
Wilson, NC 27895

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Balances	0.001 g to 400 g	$(1.16 \times 10^{-3} + 2.00 \times 10^{-6} \text{wt})$ g	Class 1 Weights
Bench Scales	0.002 lb to 10 lb	$(2.31 \times 10^{-3} + 2.80 \times 10^{-5} \text{wt})$ lb	Class F Weights
	0.01 lb to 50 lb	$(1.15 \times 10^{-2} + 2.80 \times 10^{-5} \text{wt})$ lb	
	0.02 lb to 100 lb	$(2.31 \times 10^{-2} + 2.80 \times 10^{-5} \text{wt})$ lb	
	0.05 lb to 200 lb	$(5.77 \times 10^{-2} + 2.30 \times 10^{-5} \text{wt})$ lb	
	0.1 lb to 500 lb	$(1.15 \times 10^{-1} + 2.80 \times 10^{-5} \text{wt})$ lb	
Floor Scales	0.5 lb to 2 500 lb	$(5.77 \times 10^{-1} + 2.80 \times 10^{-5} \text{wt})$ lb	Class F Weights
	1 lb to 5 000 lb	$(1.16 + 2.80 \times 10^{-5} \text{wt})$ lb	
	2 lb to 10 000 lb	$(2.31 + 2.80 \times 10^{-5} \text{wt})$ lb	
	5 lb to 20 000 lb	$(5.78 + 2.30 \times 10^{-5} \text{wt})$ lb	
Hopper Scales	0.02 lb to 100 lb	$(2.31 \times 10^{-2} + 1.15 \times 10^{-4} \text{wt})$ lb	Class F Weights
	0.1 lb to 500 lb	$(1.15 \times 10^{-1} + 2.80 \times 10^{-5} \text{wt})$ lb	
	0.5 lb to 2 500 lb	$(5.77 \times 10^{-1} + 2.80 \times 10^{-5} \text{wt})$ lb	
	1 lb to 5 000 lb	$(1.16 + 2.80 \times 10^{-5} \text{wt})$ lb	
	2 lb to 10 000 lb	$(2.31 + 2.80 \times 10^{-5} \text{wt})$ lb	
	5 lb to 20 000 lb	$(5.78 + 2.30 \times 10^{-5} \text{wt})$ lb	
	10 lb to 40 000 lb	$(11.55 + 2.30 \times 10^{-5} \text{wt})$ lb	
Truck Scales	1 000 lb to 100 000 lb	$(23.1 + 2.80 \times 10^{-5} \text{wt})$ lb	Class F Weights

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.
2. The term wt represents weight in pounds or grams (including SI multiple and submultiples units) appropriate to the uncertainty statement.