

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

***American Scale Company
7231 Covecreek Blvd
Charlotte, NC 28215***

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

***Calibration of Small, Medium and Large Weighing Devices
(As detailed in the supplement)***

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:

July 09, 2005

Issue Date:

June 23, 2009

Expiration Date:

June 22, 2011

Accreditation No.:

59302

Certificate No.:

L09-67

Page No.:

Page 1 of 2

Certificate of Accreditation: Supplement

American Scale Company
7231 Covecreek Blvd
Charlotte, NC 28215

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Scales and Balances	1 mg to 500 mg	0.025 mg	Class I Weights NIST Handbook 44
	500 mg to 5 g	0.054 mg	
	5 g to 20 g	0.074 mg	
	20 g to 50 g	0.2 mg	
	50 g to 400 g (0.11 lb to 0.88 lb)	0.06 % of reading	
	0.907 kg to 54.431 t (2 lb to 120 000 lb)	0.012 % of reading	Class F Weights NIST Handbook 44

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent method information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.