

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

***Experior Photonics
1635 Ives Avenue
Oxnard, CA 93033***

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-PLAC-IAF Communiqué dated January 2009):

***Laboratory Calibration of Fiber Optic Test Instruments
(As detailed in the supplement)***

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:

June 16, 2007

Issue Date:

June 28, 2009

Expiration Date:

June 27, 2011

Accreditation No.:

59356

Certificate No.:

L09-68

Page No.:

Page 1 of 2

Certificate of Accreditation: Supplement

Experior Photonics
1635 Ives Avenue
Oxnard, CA 93033

Accreditation is granted to this facility to perform the following calibrations:

Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Optical Power Meters 850 nm 1 300 nm 1 310 nm 1 550 nm	4 mW to 10 nW	4.1 % of reading	IEC 61315 Primary Standard Power Sensor

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.