



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory Quality System of:

***BSI Scales, Inc.
16155 W. Lincoln
New Berlin, WI 53151***

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

***Calibration of Mass, Force, and Weighing Devices
in the Laboratory and Field Sites
(As detailed in the supplement)***

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerzen
President/Operations Manager
Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:
May 15, 2005

Issue Date:
September 11, 2009

Expiration Date:
September 10, 2011

Accreditation No.:
59310

Certificate No.:
L09-87

Page No.:
Page 1 of 2



Certificate of Accreditation: Supplement

BSI Scales, Inc.
16155 W. Lincoln
New Berlin, WI 53151

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Precision balances	1 g to 420 g	$(1.15 \times 10^{-3} + 2.0 \times 10^{-6} \text{ wt}) \text{ g}$	Class I NIST Handbook 44 Test Methods Class 1 Weights
Analytical Balances and Top Loader Balances	1 g to 64 000 g	$(1.15 \times 10^{-1} + 1.14 \times 10^{-4} \text{ wt}) \text{ g}$	Class II/III NIST Handbook 44 Test Methods Class F Weights
Electronic Balances, Top Loader Type and Bench Scales	1 lb to 25 lb	$(5.75 \times 10^{-3} + 2.90 \times 10^{-5} \text{ wt}) \text{ lb}$	Class III NIST Handbook 44 Test Methods Class F Weights
	1 lb to 50 lb	$(1.15 \times 10^{-2} + 2.8 \times 10^{-5} \text{ wt}) \text{ lb}$	
	1 lb to 100 lb	$(2.31 \times 10^{-2} + 2.8 \times 10^{-5} \text{ wt}) \text{ lb}$	
Floor Scales	1 lb to 500 lb	$(1.15 \times 10^{-1} + 2.8 \times 10^{-5} \text{ wt}) \text{ lb}$	Class III NIST Handbook 44 Test Methods Class F Weights
	1 lb to 1 000 lb	$(2.31 \times 10^{-1} + 2.8 \times 10^{-5} \text{ wt}) \text{ lb}$	
	1 lb to 5 000 lb	$(1.64 + 2.0 \times 10^{-5} \text{ wt}) \text{ lb}$	
	2 lb to 10 000 lb	$(2.31 + 2.8 \times 10^{-5} \text{ wt}) \text{ lb}$	
	5 lb to 20 000 lb	$(5.78 + 2.3 \times 10^{-5} \text{ wt}) \text{ lb}$	
Vehicle, Axle, Livestock, Crane, Hopper and Rail Scales	10 lb to 50 000 lb	$(5.78 + 4.8 \times 10^{-5} \text{ wt}) \text{ lb}$	Class III & IIII NIST Handbook 44 Test Methods Class F Weights
	20 lb to 100 000 lb	$(23.1 + 2.8 \times 10^{-5} \text{ wt}) \text{ lb}$	
	20 lb to 200 000 lb	$(32.66 + 3.7 \times 10^{-5} \text{ wt}) \text{ lb}$	

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method.
2. The term wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.