



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory of:

***Allied Scale Inc.
13654 S. Kenton Avenue
Crestwood, IL 60445***

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025: 2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

***Field Calibration of Customer-Owned Weigh Scales and Equipment
(As detailed in the supplement)***

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
26555 Evergreen, Suite 1325
Southfield, Michigan 48076

Initial Accreditation Date:

July 23, 2003

Issue Date:

October 02, 2009

Expiration Date:

October 01, 2011

Accreditation No.:

25830

Certificate No.:

L09-98

Page No.:

Page 1 of 2



Certificate of Accreditation: Supplement

Allied Scale Inc.
13654 S. Kenton Avenue
Crestwood, IL 60445

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	REMARKS
Weigh Scales	2 mg to 1 kg	$(0.012 + 1.05 \times 10^{-4} \text{ wt}) \text{ g}$	Class F Weights
	0.001 lb to 25 lb	$(0.012 + 1.5 \times 10^{-5} \text{ wt}) \text{ lb}$	Class F Weights
	50 lb to 1 000 lb	$(0.12 + 5.1 \times 10^{-5} \text{ wt}) \text{ lb}$	Class F Weights
	1 000 lb to 26 000 lb	29 lb	Class F Weights

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method.
2. The term wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.