



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc., has assessed the Laboratory Quality System of:*

***Applied Scale Technology  
3012 Ambrose Ave.  
Nashville, TN 37207***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

***ISO/IEC 17025: 2005***

*This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):*

***Calibration of Weigh Scales and Analytical Balances  
(As detailed in the supplement)***

*Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.*

For PJLA:

*The validity of this certificate is mandated through ongoing surveillance.*

Tracy Szerszen  
President/Operations Manager

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
26555 Evergreen, Suite 1325  
Southfield, Michigan 48076

*Initial Accreditation Date:*

October 01, 2003

*Issue Date:*

October 17, 2009

*Expiration Date:*

October 16, 2011

*Accreditation No.:*

24400

*Certificate No.:*

L09-99

*Page No.:*

Page 1 of 2



# Certificate of Accreditation: Supplement

**Applied Scale Technology**  
3012 Ambrose  
Nashville, TN 37207

*Accreditation is granted to this facility to perform the following calibrations:*

## Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	REMARKS
Bench Scales	1 lb to 300 lb	$(5.78 \times 10^{-2} + 3.21 \times 10^{-5} \text{wt}) \text{ lb}$	Class F
Floor Scales	1 lb to 5 000 lb	$(1.16 + 2.73 \times 10^{-5} \text{wt}) \text{ lb}$	Class F
Truck Scales	20 lb to 120 000 lb	$(23.1 + 3.2 \times 10^{-5} \text{wt}) \text{ lb}$	Class F
Lab Balances	10 mg to 300 g	$(2.0 \times 10^{-4} + 2.57 \times 10^{-6} \text{wt}) \text{ g}$	Class 1
	301 g to 200 kg	$(1.0 \times 10^{-2} + 5.72 \times 10^{-6} \text{wt}) \text{ g}$	Class 2

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method.
2. The term wt represents weight in pounds or grams (including SI multiple and submultiples units) appropriate to the uncertainty statement.