



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Aerospace Metrology & Electromechanical Calibration Ltd.
Met Cal House, Fisher Street, Newcastle-Upon-Tyne NE6 4LT, UK

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Optical Measurements
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Initial Accreditation Date:

June 19, 2019

Issue Date:

June 19, 2019

Expiration Date:

March 26, 2020

Tracy Szerszen
President/Operations Manager

Accreditation No.:

106685

Certificate No.:

L19-299

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Aerospace Metrology & Electromechanical Calibration Ltd.

Met-Cal House, Fisher Street, Newcastle-upon-Tyne, NE6 4LT, UK
Contact Name: Stephan Oxborough Phone: 191-262-2266

Accreditation is granted to the facility to perform the following calibrations:

Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Gloss ^F Measured at fixed geometries of	20° Mirror	24 SGU	Gloss Standards
	20° High Gloss	0.54 SGU	
	20° Semi Gloss	0.66 SGU	
	60° Mirror	22 SGU	
	60° High Gloss	0.54 SGU	
	60° Semi Gloss	0.66 SGU	
	85° Mirror	1.1 SGU	
	85° High Gloss	0.65 SGU	
	85° Semi Gloss	0.76 SGU	
	At geometries of 20°, 60°, 85° 0 GU to 100 GU		0.8 GU
101 GU to 2 000 GU		1.1 GU	

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.