



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Global Calibration, LLC

1507-11 Smithtown Avenue, Bohemia, NY 11716

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Electrical, Dimensional Calibration, Mechanical, Time & Frequency, and
Thermodynamic
(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

November 01, 2017

Issue Date:

November 26, 2023

Expiration Date:

November 26, 2025

Accreditation No.:

85580

Certificate No.:

L23-854

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a
continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjllabs.com*



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure DC Voltage ^{FO}	Up to 330 mV	17 µV + 39.3 µV/V	Fluke 5502A GC 146
	0.33 mV to 33 V	0.51 µV + 0.18 µV/V	
	3.3 V to 33 V	5.9 mV + 9.9 µV/V	
	33 V to 330 V	42 mV + 17 µV/V	
	330 V to 1 020 V	150 mV + 17 µV/V	
Equipment to Output DC Voltage ^{FO}	Up to 100 mV	1.9 µV + 1.9 nV/mV	HP 3458A/Opt 2 GC 174
	100 mV to 1 V	15 µV + 1.9 nV/mV	
	1V to 10 V	52 µV + 4.2 µV/V	
	10 V to 100 V	72 µV + 6.6 µV/V	
	100 V to 1 000 V	5.5 mV + 6.5 µV/V	
Equipment to Measure AC Voltage At the Listed Frequencies ^{FO}			Fluke 5502A GC 146
10 Hz to 45 Hz	Up to 33 mV	0.18 mV + 1.3 mV/V	
45 Hz to 10 kHz	Up to 33 mV	58 µV + 1.2 mV/V	
10 kHz to 20 kHz	Up to 33 mV	74 µV + 2.4 mV/V	
20 kHz to 50 kHz	Up to 33 mV	55 µV + 2.3 mV/V	
50 kHz to 100 kHz	Up to 33 mV	62 µV + 3.7 mV/V	
100 kHz to 500 kHz	Up to 33 mV	0.43 mV + 0.82 mV/V	
Equipment to Measure AC Voltage At the Listed Frequencies ^{FO}			
10 Hz to 45 Hz	33 mV to 330 mV	0.13 mV + 2.7 mV/V	
45 Hz to 10 kHz	33 mV to 330 mV	49 µV + 0.53 mV/V	
10 kHz to 20 kHz	33 mV to 330 mV	25 µV + 1.7 mV/V	
20 kHz to 50 kHz	33 mV to 330 mV	47 µV + 1.9 mV/V	
50 kHz to 100 kHz	33 mV to 330 mV	200 µV + 2.9 mV/V	
100 kHz to 500 kHz	33 mV to 330 mV	0.41 mV + 0.82 mV/V	
Equipment to Measure AC Voltage At the Listed Frequencies ^{FO}			
10 Hz to 45 Hz	0.33 V to 3.3 V	0.14 mV + 0.33 mV/V	
45 Hz to 10 kHz	0.33 V to 3.3 V	0.14 mV + 0.34 mV/V	
10 kHz to 20 kHz	0.33 V to 3.3 V	0.18 mV + 1.1 mV/V	
20 kHz to 50 kHz	0.33 V to 3.3 V	0.36 mV + 1.6 mV/V	
50 kHz to 100 kHz	0.33 V to 3.3 V	1.9 mV + 2.9 mV/V	
100 kHz to 500 kHz	0.33 V to 3.3 V	3.7 mV + 6.2 mV/V	



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Voltage At the Listed Frequencies ^{FO}			Fluke 5502A GC 146
10 Hz to 45 Hz	3.3 V to 33 V	3.5 mV + 1.7 mV/V	
45 Hz to 10 kHz	3.3 V to 33 V	4.3 mV + 0.38 mV/V	
10 kHz to 20 kHz	3.3 V to 33 V	20 mV + .92 mV	
20 kHz to 50 kHz	3.3 V to 33 V	6.6 mV + 2.2 mV/V	
50 kHz to 100 kHz	3.3 V to 33 V	20 mV + 2.9 mV/V	
Equipment to Measure AC Voltage At the Listed Frequencies ^{FO}			Fluke 5502A GC 146
45 Hz to 1 kHz	33 V to 330 V	20 mV + 0.92 mV/V	
1 kHz to 10 kHz	33 V to 330 V	20 mV + 0.92 mV/V	
10 kHz to 20 kHz	33 V to 330 V	0.01 V + 5.9m V/V	
Equipment to Measure AC Voltage At the Listed Frequencies ^{FO}			HP 3458A/Opt 2 GC 174
45 Hz to 1 kHz	33 V to 330 V	20 mV + 0.92 mV/V	
1 kHz to 10 kHz	33 V to 330 V	20 mV + 0.92 mV/V	
10 kHz to 20 kHz	33 V to 330 V	0.01 V + 5.9 m V/V	
Equipment to Output AC Voltage At the Listed Frequencies ^{FO}			
1 Hz to 40 Hz	Up to 10 mV	0.1 mV + 0.85 μV/mV	
40 Hz to 1 kHz	Up to 10 mV	17 μV + 0.23 μV/mV	
1 kHz to 20kHz	Up to 10 mV	10 μV + 0.32 μV/mV	
20kHz to 50 kHz	Up to 10 mV	1.4 μV + 1.2 μV/mV	
50 kHz to 100 kHz	Up to 10 mV	1.7 μV + 5.8 μV/mV	
100 kHz to 300 kHz	Up to 10 mV	4.2 μV + 46 μV/mV	
Equipment to Output AC Voltage At the Listed Frequencies ^{FO}			
1 Hz to 40 Hz	10 mV to 100 mV	4.1 μV + 0.13 μV/mV	
40 Hz to 1 kHz	10 mV to 100 mV	13 μV + 48 nV/mV	
1 kHz to 20kHz	10 mV to 100 mV	13 μV + 98 nV/mV	
20kHz to 50 kHz	10 mV to 100 mV	9.3 μV + 0.12 μV/mV	
50 kHz to 100 kHz	10 mV to 100 mV	7.6 μV + 1.1 μV/mV	
100 kHz to 300 kHz	10 mV to 100 mV	16 μV + 3.4 μV/mV	
300 kHz to 1 MHz	10 mV to 100 mV	82 μV + 11 μV/mV	
1MHz to 2 MHz	10 mV to 100 mV	61 μV + 17 μV/mV	



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output AC Voltage At the Listed Frequencies ^{FO}			HP 3458A/Opt 2 GC 174
1 Hz to 40 Hz	0.1 V to 1 V	92 μV + 60 nV/mV	
40 Hz to 1 kHz	0.1 V to 1 V	64 μV + 71 nV/mV	
1 kHz to 20kHz	0.1 V to 1 V	45 μV + 0.16 nV/mV	
20kHz to 50 kHz	0.1 V to 1 V	63 μV + 0.15 μV/mV	
50 kHz to 100 kHz	0.1 V to 1 V	0.39 mV + 0.88 μV/mV	
100 kHz to 300 kHz	0.1 V to 1 V	0.13 mV + 3.7 μV/mV	
300 kHz to 1 MHz	0.1 V to 1 V	0.78 mV + 11 μV/mV	
1MHz to 2 MHz	0.1 V to 1 V	0.69 mV + 17 μV/mV	
Equipment to Output AC Voltage At the Listed Frequencies ^{FO}			
1 Hz to 40 Hz	1 V to 10 V	0.73 mV + 77 mV/mV	
40 Hz to 1 kHz	1 V to 10 V	0.46 mV + 81 μV/mV	
1 kHz to 20kHz	1 V to 10 V	0.22 mV + 18 μV/mV	
20kHz to 50 kHz	1 V to 10 V	0.63 mV + 0.15 mV/mV	
50 kHz to 100 kHz	1 V to 10 V	0.8 mV + 1.1 mV/V	
100 kHz to 300 kHz	1 V to 10 V	1.3 mV + 3.7 mV/V	
300 kHz to 1 MHz	1 V to 10 V	7.8 mV + 11 mV/V	
1MHz to 2 MHz	1 V to 10 V	6.9mV + 17 mV/V	
Equipment to Output AC Voltage At the Listed Frequencies ^{FO}			
1 Hz to 40 Hz	10 V to 100 V	4.5 mV + 0.24 mV/V	
40 Hz to 1 kHz	10 V to 100 V	4.1 mV + 0.22 mV/V	
1 kHz to 20kHz	10 V to 100 V	4 mV + 0.23 mV/V	
20kHz to 50 kHz	10 V to 100 V	6 mV + 0.21 mV/V	
50 kHz to 100 kHz	10 V to 100 V	5.5 mV + 1.5 mV/V	
100 kHz to 300 kHz	10 V to 100 V	13 mV + 4.8 mV/V	
300 kHz to 1 MHz	10 V to 100 V	78 mV +11 mV/V	
Equipment to Output AC Voltage At the Listed Frequencies ^{FO}			
1 Hz to 40 Hz	100 V to 1 000 V	81 mV + 0.43 mV/V	
40 Hz to 1 kHz	100 V to 1 000 V	220 mV + 0.29 mV/V	
1 kHz to 20kHz	100 V to 1 000 V	190 mV + 0.71 mV/V	
20kHz to 50 kHz	100 V to 1 000 V	62 mV + 0.65 mV/V	
50 kHz to 100 kHz	100 V to 1 000 V	40 mV + 3.5 mV/V	



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure Resistance ^{FO}	2 Ω to 10.9 Ω	0.45 Ω + 5.5 $\mu\Omega/\Omega$	Fluke 5502A GC 146
	11 Ω to 32.9 Ω	0.45 Ω + 6.2 $\mu\Omega/\Omega$	
	33 Ω to 109 Ω	0.45 Ω + 5.6 $\mu\Omega/\Omega$	
	110 Ω to 329 Ω	0.45 Ω + 8.3 $\mu\Omega/\Omega$	
	330 Ω to 1 090 Ω	0.48 Ω + 0.27 m Ω /k Ω	
	1.1 K Ω to 3.29 K Ω	0.45 Ω + 54 Ω /k Ω	
	3.3 K Ω to 10.9 K Ω	0.89 Ω + 95 m Ω /k Ω	
	11 K Ω to 32.9 K Ω	1.5 Ω + 88 m Ω /k Ω	
	33 K Ω to 109 K Ω	17 Ω + 86 m Ω /k Ω	
	110 K Ω to 329 K Ω	15 Ω + 0.12 Ω /k Ω	
	330 K Ω to 1 090 K Ω	160 Ω + 0.12 Ω /k Ω	
	1.1 M Ω to 3.29 M Ω	51 Ω + 0.19 Ω /k Ω	
	3.3 M Ω to 10.9 M Ω	1.1 k Ω + 0.67 Ω /k Ω	
	11 M Ω to 32.9 M Ω	12 k Ω + 0.96 Ω /k Ω	
	33 M Ω to 109 M Ω	190 k Ω + 5.7 Ω /k Ω	
	110 M Ω to 330 M Ω	690 k Ω + 4.4 Ω /k Ω	
Equipment to Source Resistance ^{FO}	Up to 10 Ω	1.9 m Ω + 1.4 $\mu\Omega/\Omega$	HP 3458A Opt.2 (4 wire Ω) GC 174
	10 Ω to 100 Ω	3.89 m Ω + 2.5 $\mu\Omega/\Omega$	
	100 Ω to 1 k Ω	0.29 Ω + 1.2 m Ω /k Ω	HP 3458A Opt.2 (2 wire Ω) GC 174
	1 k Ω to 10 k Ω	3.2 Ω + 3.9 m Ω /k Ω	
	10 k Ω to 100 k Ω	3.2 Ω + 3.9 m Ω /k Ω	
	100 k Ω to 1 M Ω	3.1 Ω + 26 Ω /M Ω	
	1 M Ω to 10 M Ω	120 Ω + 60 Ω /M Ω	
	10 M Ω to 100 M Ω	12 K Ω + 51 k Ω /M Ω	
	100 M Ω to 1 G Ω	1.7 M Ω + 4.4 k Ω /M Ω	



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure DC Current ^{FO}	0.02 mA to 3.2 mA	0.6 uA + 0.65 uA/mA	Fluke 5502A GC 146
	3.3 mA to 32 mA	6 uA + 0.038 uA/mA	
	33 mA to 320 mA	74 uA + 0.77 uA/mA	
	320 mA to 2.1 A	880 uA + 0.16 mA/mA	
	2.2 A to 11 A	4.4 mA 0.69 uA/mA	
Equipment to Output DC Current ^{FO}	1 nA to 100 nA	0.8 nA	HP 3458A/Opt 2 GC 174
	0.1 µA to 1 µA	0.01 + 8.1 nA/uA	
	1 µA to 10 µA	0.8 nA + 9.3 pA/uA	
	10 uA to 100 µA	0.56 pA + 33 pA/uA	
	0.1 mA to 1 mA	7.7 nA + 28 nA/mA	
	1 mA to 10 mA	7.6 nA + 28 nA/mA	
	10 mA to 100mA	0.59 µA + 46 nA/mA	
	0.1 mA to 1mA	12 µA + 13 µA/mA	
Equipment to Measure AC Current At the Listed Frequencies ^{FO}			Fluke 5502A GC 146
10 Hz to 20 Hz	Up to 330 uA	0.95 µA + 2.5 µA/mV	
20 Hz to 45 Hz	Up to 330 uA	1.1 µA + 3 µA/mA	
45 Hz to 1 kHz	Up to 330 uA	0.94 µA + 4.7 µA/mA	
1 kHz to 5 kHz	Up to 330 uA	0.89 µA + 6.1 µA/mA	
5 kHz to 10 kHz	Up to 330 uA	0.82 µA + 17 µA/mA	
Equipment to Measure AC Current At the Listed Frequencies ^{FO}			
10 Hz to 20 Hz	0.33 mA to 3.3 mA	1.6 uA + 2 µA/mA	
20 Hz to 45 Hz	0.33 mA to 3.3 mA	1.1 µA + 2.3 µA/mA	
45 Hz to 1 kHz	0.33 mA to 3.3 mA	1.2 µA + 1.1 µA/mA	
1 kHz to 5 kHz	0.33 mA to 3.3 mA	1.4 µA + 5 µA/mA	
5 kHz to 10 kHz	0.33 mA to 3.3 mA	1.3 µA + 13 µA/mA	



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Current At the Listed Frequencies ^{FO}			Fluke 5502A GC 146
10 Hz to 20 Hz	3.3 mA to 33 mA	7.1 µA + 2.6 µA/mA	
20 Hz to 45 Hz	3.3 mA to 33 mA	6.7 µA + 2.4 µA/mA	
45 Hz to 1 kHz	3.3 mA to 33 mA	7.7 µA + 1.2 µA/mA	
1 kHz to 5 kHz	3.3 mA to 33 mA	25 µA + 3 µA/mA	
5 kHz to 10 kHz	3.3 mA to 33 mA	5.8 µA + 9.4 µA/mA	
Equipment to Measure AC Current At the Listed Frequencies ^{FO}			
10 Hz to 20 Hz	33 mA to 330 mA	0.54 mA + 1.7 µA/mA	
20 Hz to 45 Hz	33 mA to 330 mA	0.55 mA + 1.3 µA/mA	
45 Hz to 1 kHz	33 mA to 330 mA	570 µA + 0.48 µA/mA	
1 kHz to 5 kHz	33 mA to 330 mA	0.89 µA + 6.1 µA/mA	
5 kHz to 10 kHz	33 mA to 330 mA	55 µA + 7.8 µA/mA	
Equipment to Measure AC Current At the Listed Frequencies ^{FO}			
10 Hz to 45 Hz	0.33 A to 2.1 A	7.5 mA + 0.55 µA/mA	
45 Hz to 1 kHz	0.33 A to 2.1 A	7.1 mA + 0.51 µA/mA	
1 kHz to 5 kHz	0.33 A to 2.1 A	3.5 mA + 14 µA/mA	
Equipment to Measure AC Current At the Listed Frequencies ^{FO}			
45 Hz to 65 Hz	2.2 A to 11 A	57 mA + 0.47 uA/mA	
65 Hz to 500 Hz	2.2 A to 11 A	58 mA + 0.3 uA/mA	
500 kHz to 1 kHz	2.2 A to 11 A	53 mA + 3.2 uA/mA	



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Capacitance – Source ^{FO}	0.33 nF to 0.49 nF	12 pF + 0.005 9 pF/pF	Fluke 5502A GC 146
	0.5 nF to 1.09 nF	1.7 pF + 16.2 nF/uF	
	1.1 nF to 3.29 nF	12 pF + 6 nF/uF	
	3.3 nF to 10.09 nF	12 pF + 5.9 nF/uF	
	11 nF to 32.9 nF	0.22 nF + 0.16 nF/uF	
	33 nF to 109 nF	0.59 nF + 1.4 nF/uF	
	110 nF to 329 nF	3.8 nF + 7.1 nF/uF	
	0.33 uF to 10.09 uF	80 nF + 0.092 nF/uF	
	11 uF to 32.9 uF	0.15 nF + 5.9nF/uF	
	33 uF to 100.9 uF	170 nF + 5.3nF/uF	
	110 uF to 329 uF	350 nF + 8.2nF/uF	
	330 uF to 1.1 mF	3.3 uF	
Oscilloscopes – DC Voltage ^{FO}	500 μ V to 130 V 1 M Ω	130 μ V + 0.15 μ V/mV	Fluke 5800A GC 120
	200 μ V to 6.6 V 50 Ω	46 μ V + 2.9 μ V/mV	
Leveled Sine Wave Reference 50 kHz ^{FO}	5 mV to 5.5 Vp-p	0.34 mV + 24 mV/V (p-p)	
50 kHz to 100 MHz	5 mV to 5.5 Vp-p	0.035 mV + 42 mV/V (p-p)	
100 MHz to 100 MHz	5 mV to 5.5 Vp-p	0.35 mV + 65 mV/V (p-p)	
50 kHz to 300 MHz	5 mV to 5.5 Vp-p	0.035 mV + 65 mV/V (p-p)	
300 MHz to 500 MHz	5 mV to 5.5 Vp-p	0.35 mV + 65 mV/V (p-p)	
500 MHz to 600 MHz	5 mV to 5.5 Vp-p	0.35 mV + 71 mV/V (p-p)	



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Square wave 50 Ω Load (> 1 kHz) ^{FO}	1 mV to 24.9 mV	44 μ V + 3.1 μ V/mV	Fluke 5800A GC 120
	25 mV to 109 mV	46 μ V + 2.9 μ V/mV	
	110 mV to 2.1 V	57 μ V + 2.9 μ V/mV	
	2.2 V to 6.6 V	110 μ V + 2.9 μ V/mV	
Square wave 1 M Ω Load (10 Hz to 10 kHz) ^{FO}	1 mV to 24.9 mV	46 μ V + 0.59 μ V/mV	
	25 mV to 109 mV	49 μ V + 0.57 μ V/mV	
	110 mV to 2.1 V	74 μ V + 0.36 μ V/mV	
	2.2 V to 10.9 V	530 μ V + 0.58 μ V/mV	
	11 V to 130 V	20 mV + 1.5 mV/V	
Time Marker ^{FO}	2 ns to 5 ns	0.58 ps/ns	
	10 ns	6 ps	
	20 ns to 50 ns	0.58 ps/ns	
	100 ns to 20 ms	77 ps/ns	
	50 ms to 5 s	580 ps/ms	
Rise Time ^{FO}	50 Ω load	289 ps	
Input Resistance Measurement ^{FO}	40 Ω to 60 Ω	0.52 m Ω + 1.3 m Ω / Ω	Fluke 5800A Oscilloscope Calibrator Fluke user manual DUT Manual GC 120
	500 k Ω to 1.5 M Ω	22 Ω + 1.2 k Ω /M Ω	
Input Capacitance Measurement ^{FO}	5 pF to 50 pF	0.58 pF + 0.058 pF/pF	

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Thermocouple Simulation Type N ^{FO}	-200 $^{\circ}$ C to -100 $^{\circ}$ C	0.47 $^{\circ}$ C	Fluke 5502A GC 106
	-100 $^{\circ}$ C to -25 $^{\circ}$ C	0.32 $^{\circ}$ C	
	-25 $^{\circ}$ C to 410 $^{\circ}$ C	0.3 $^{\circ}$ C	
	410 $^{\circ}$ C to 1 300 $^{\circ}$ C	0.36 $^{\circ}$ C	
Thermocouple Simulation Type K ^{FO}	-200 $^{\circ}$ C to -100 $^{\circ}$ C	0.41 $^{\circ}$ C	
	-100 $^{\circ}$ C to -25 $^{\circ}$ C	0.3 $^{\circ}$ C	
	-25 $^{\circ}$ C to 120 $^{\circ}$ C	0.29 $^{\circ}$ C	
	120 $^{\circ}$ C to 1 000 $^{\circ}$ C	0.35 $^{\circ}$ C	
	1 000 $^{\circ}$ C to 1 372 $^{\circ}$ C	0.46 $^{\circ}$ C	



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Thermocouple Simulation Type R ^{FO}	Up to 250 °C	0.62 °C	Fluke 5502A GC 106
	250 °C to 400 °C	0.43 °C	
	400 °C to 1 000 °C	0.41 °C	
	1 000 °C to 1 767 °C	0.46 °C	
Thermocouple Simulation Type S ^{FO}	Up to 250 °C	0.53 °C	
	250 °C to 1 000 °C	0.43 °C	
	1 000 °C to 1 400 °C	0.44 °C	
	1 400 °C to 1 767 °C	0.52 °C	
Thermocouple Simulation Type T ^{FO}	-250 °C to -150 °C	0.67 °C	
	-150 °C to 0 °C	0.34 °C	
	Up to 120 °C	0.29 °C	
	120 °C to 400 °C	0.27 °C	
Type K Thermocouples ^F	50 °C to 660 °C	0.13 + (0.000 608T) °C	Dry Well, SPRT, & HP 3458 GC 148
Type N Thermocouples ^F	50 °C to 660 °C	0.19 + (0.000 785T) °C	
Type T Thermocouples ^F	50 °C to 400 °C	0.13 + (0.000 737T) °C	
Temperature Measuring Systems ^F	-30 °C to 420 °C	0.015 °C	Fluke Dry Block & SPRT GC 128
Temperature Blocks ^F	-30 °C to 140 °C	0.015 °C	Fluke 1521 & SPRT GC 115
Temperature Chambers, Ovens, incubators ^O	-30 °C to 200 °C	0.93 °C	"T" Thermocouples & Readout GC 137



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pressure Gauge ^{FO}	Up to 10 inWC	(0.006 3 + 0.000 012 9 P) inWC	Ashcroft ATE & AQS1-10 GC 105B
	Up to 200 inWC	0.13 inWC	Ashcroft ATE & HQS1-200 GC 105B
	Up to 20 inWC	0.000 94 + (0.000 051 3 .9 P) inWC	Fluke 7250LP GC 105B
	20 to 60 inWC	0.000 32 + (000 095 9 P) inWC	
	Up to 250 Psig	(0.005 7 + 0.001 3 P) Psig	Druck DPI 310A GC 105B
	250 Psig to 1 250 Psig	(0.005 5 + 0.001 2 P) Psig	
	1 250 psig to 6 250 psig	(0.009 1 + 0.001 1 P) psig	
	6 250 psig to 12 500 psig	(0.078 + 0.001 4 P) psig	
Torque Analyzers ^F	5 lbf·in to 100 lbf·in	0.006 9 + (0.000 795 TQ) lbf·in	Torque Wheel 5" & Class F weights GC 117
	10 lbf·in to 500 lbf·in	0.11 + (0.000 55 TQ) lbf·in	Torque Wheel 10" & Class F weights GC 117
	360 lbf·in to 2 400 lbf·in	0.38 + (0.000 09 TQ) lbf·in	Torque Arm 36" & Class F weights GC 117
	40 lbf·in to 1 000 lbf·ft	0.087 + (0.000 097 TQ) lbf·ft	Torque Arm 48" & Class F weights GC 117
Torque Wrenches ^{FO}	10 lbf·in to 100 lbf·in	0.25 + (0.004 04 TQ) lbf·in	Torque Analyzer/Tester GC 118
	120 lbf·in to 600 lbf·in	2.8 + (0.000 717 TQ) lbf·in	
	50 lbf·ft to 250 lbf·ft	0.58 + (0.001 22 TQ) lbf·ft	
Gas Flow Meters ^{FO}	2 sccm to 20 sccm	0.028 sccm + 0.001 sccm/sccm	Fluke-DH Molbox RFM-M Molboc 2E2 Molboc 1SLM Molboc 5E3 Molboc 10SLM GC 121
	20 sccm to 200 sccm	0.16 sccm + 0.001 5 sccm/sccm	
	10 sccm to 100 sccm	0.13 sccm + 0.001 sccm/sccm	
	100 sccm to 1 000 sccm	0.47 sccm + 0.018 sccm/sccm	
	50 sccm to 500 sccm	0.59 sccm + 0.1 sccm/sccm	
	500 sccm to 5 000 sccm	1.2 sccm + 0.001 9 sccm/sccm	
	100 sccm to 1 000 sccm	1.3 sccm + 0.000 71 sccm/sccm	
	1 000 sccm to 10 000 sccm	4.5 sccm + 0.019 sccm/sccm	
Liquid Flow Meters ^{FO}	0.1 Gpm to 15 Gpm	0.002 9 Gpm + 0.000 837 Gpm/Gpm	Coriolis Flow meters GC 109
	10 Gpm to 300 Gpm	0.038 Gpm + 0.001 67 Gpm/Gpm	



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

Time & Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Stopwatches, Timers ^{FO}	Up to 24 hrs	0.059 s/day	Vibrograf Timometer GC 138
Frequency ^{FO}	2 Hz to 100 Hz	0.58 Hz + 0.14 μ Hz/Hz	Fluke 5500 fluke user manual DUT Manual GC 146
	100 Hz to 100 KHz	0.58 Hz + 59 μ Hz/Hz	
	100 KHz to 2 MHz	0.0 Hz + 1 mHz/Hz	

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Calipers ^{FO}	Up to 12 in	290 μ in + 2.9 μ in/in	Gage Blocks and Plain Ring Gage GC 124
	12 in to 24 in	370 μ in + 1.7 μ in/in	
	24 in to 60 in	31 μ in + 0.5 μ in/in	
Micrometers Outside / Inside	Up to 1 in	56 μ in + 1.9 μ in/in	Gage Blocks GC 164
	1 in to 6 in	51 μ in + 3.33 μ in/in	
Indicators	Up to 1 in	49 μ in + 9 μ in/in	Gage Blocks GC 129
Thread Plug Gages ^F Pitch Diameter	Up to 4 in	41 μ in + 17 μ in/in	Super micrometer, Thread wires and Gage Blocks GC 162
Thread Plug Gages ^F Major Diameter	Up to 4 in	56 μ in + 2.3 μ in/in	Super micrometer and Gage Blocks GC 162
Gage Block GGG3 & B89 grade 0 ^F	0.1 in to 1 in	2 μ in + 3.7 μ in/L	Master Gage Blocks And Comparator GC113
	2 mm to 25.4 mm	55 nm + 5.75 nm/L	
	1 in to 7 in	1.2 μ in + 5.3 μ in/L	
	25 mm to 180 mm	0.17 μ m + 4.5 nm/L	



Certificate of Accreditation: Supplement

Global Calibration, LLC

1507-11 Smithtown, Bohemia, NY 11716

Contact Name: Mr. Charles Mahoney Phone: 631-750-5663

Accreditation is granted to the facility to perform the following calibrations:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations.
5. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
6. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
7. The term T represents temperature in degrees Celsius as appropriate to the uncertainty statement.
8. The term P represents pressure in pounds per square inch gage or inches of water column as appropriate to the uncertainty statement.