



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Accelerated Analytical Labs, Inc.
9075 West Heather Avenue, Milwaukee, WI 53224

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017
& Meets the Requirements of the AOAC International Guidelines for Laboratories Performing Microbiological and Chemical Analyses of Food and Pharmaceutical 2018

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Microbiological and Chemical Testing
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

May 06, 2013

Issue Date:

September 11, 2023

Expiration Date:

December 31, 2025

Revision Date:

September 09, 2024

Accreditation No.:

76006

Certificate No.:

L23-678-R1

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Accelerated Analytical Labs, Inc.

9075 West Heather Avenue, Milwaukee, WI 53224

Contact Name: Nathan Greenleaf Phone: (414) 867-7007 Ext 113

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Microbiological ^F	Food Products, Feeds, Cosmetics, Nonwovens, Lip Balms, Drugs and Food additives, OTC, and Unknown Biological Samples	Heterotrophic Plate Count	SM 9215	1 CFU/g or 1 CFU/mL
		Microbial Characterization, Identification, and Strain Typing	USP <1113>	Qualitative
		<i>Burkholderia cepacia</i> Complex	USP <60>	Presence/Absence
		Examination of Nonsterile Products: Microbial Enumeration	USP <61>	1 CFU/g or 1 CFU/mL
		Examination of Nonsterile Products: Specified Microorganisms	USP <62>	Presence/Absence
		Antimicrobial Effectiveness	USP <51>	NA or Pass/Fail
		Chemical ^F	Aqueous, Non-Aqueous Liquids and Solids	Heavy Metals: As Cd Na Pb Hg
Chromatography, GC Identifiable Compounds	USP <621> Assay, Impurities, Chromatographic Purity			Qualitative / Quantitative 100 ppm (Compound Specific)
Chromatography, HPLC Identifiable Compounds				
Heavy Metals: As Pb Hg	USP <233>			LOD = 1 ppm
Melting Range	USP <741>			25 °C to 300 °C
pH	USP <791>			1 pH to 12 pH
Residual Solvents Class 1 and 2: Benzene Carbon tetrachloride 1,2-Dichloroethane 1,1-Dichloroethene 1,1,1-Trichloroethane Acetonitrile Chlorobenzene Chloroform Cumene Cyclohexane 1,2-Dichloroethene	USP <467>			LOD = 2 ppm (Compound Specific)



Certificate of Accreditation: Supplement

Accelerated Analytical Labs, Inc.

9075 West Heather Avenue, Milwaukee, WI 53224

Contact Name: Nathan Greenleaf Phone: (414) 867-7007 Ext 113

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Chemical ^F	Aqueous, Non-Aqueous Liquids and Solids	Residual Solvents Class 1 and 2: 1,2-Dimethoxyethane N,N-Dimethylacetamide N,N-Dimethylformamide 1,4-Dioxane 2-Ethoxyethanol Ethylene glycol Formamide Hexane Methanol 2-Methoxyethanol Methylbutylketone Methylcyclohexane Methylene chloride Methylisobutylketone N-Methylpyrrolidone Nitromethane Pyridine Sulfolane Tetrahydrofuran Tetralin Toluene Trichloroethylene Xylene	USP <467>	LOD = 2 ppm (Compound Specific)
		Specific Gravity	USP <841>	LOD = 0.5 to 2
		Spectrometric Identification ID, Spectrometric Identification IR	USP <197>	Qualitative, 4 000 cm ⁻¹ to 200 cm ⁻¹
		Spectrometric Identification ID, Spectrometric Identification UV		Qualitative, 200 nm to 650 nm
		Spectrophotometry IR Identifiable Compounds	USP <854> Spectrophotometry IR	Qualitative / Quantitative: 100 ppm LOD (Compound Specific)
		Spectrophotometry, UV Identifiable compounds	USP <857> Spectrophotometry UV	
		Titrimetry, Indicator and Potentiometric	USP <541>	0.01 % LOD, compound specific
		Viscosity	USP <912>	0 cP to 1 000 000 cP
		Water Determination	USP <921>	LOD = 0.01 %



Certificate of Accreditation: Supplement

Accelerated Analytical Labs, Inc.

9075 West Heather Avenue, Milwaukee, WI 53224

Contact Name: Nathan Greenleaf Phone: (414) 867-7007 Ext 113

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Chemical ^F	Water	TOC	USP <643>	LOD = 0.05 ppm
		Conductivity	USP <645>	LOD = 1 µmhos/cm

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer ^F would mean that the laboratory performs this testing at its fixed location.

